

Certificate of Conformity

NucleoCounter[®] SCC-400[™]

Name of product: NucleoCounter[®] SCC-400[™]
Type: SCC-400[™]
Other identifying data: Part no. 900-0400

We declare under our sole responsibility that the products, to which this declaration relates, are in conformity with the Council Directives on the approximation of the laws of the EEC Member States relating to the following:

Directives	Electromagnetic Compatibility (2004/108/EC). EMC Low Voltage Directive (2006/95/EC) LVD Machinery Directive (2006/42/EC) MD
Standards EMC	EN 61326-1:2006 Electrical equipment for laboratory use EN 61000-3-2: 2006 Limits for harmonic current emissions EN 61000-3-3: 1995 +A1 +A2:2005 Limitation of voltage changes, voltage fluctuations and flicker
Standards LVD	EN 61010-1:2001 Safety requirements for electrical equipment for measurement, control, and laboratory use
Standards MD	EN 12100-1: 2003 Safety of machinery; Basic concepts, general principles for design — Part 1 EN 12100-2: 2003 Safety of machinery; Basic concepts, general principles for design — Part 2: EN ISO 14121-1:2007 Safety of machinery; risk assessment

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Date: 2014-12-15


Signed:



Name: Louise L. B. Christoffersen
Position: QA/QC Manager

Name and address of manufacturer:

ChemoMetec A/S
Gydevang 43
DK-3450 Allerod
Denmark



chemometec

ChemoMetec A/S
Gydevang 43
DK-3450 Allerod
Denmark
Phone(+45) 48 13 10 20
Fax (+45) 48 13 10 21
Mail info@chemometec.com
Web www.chemometec.com