Certificate of Analysis for lot no. 0123-211

NC-200 PQ Kit 4

Kit Part no.: 912-0013
Lot no.: 0123-211
Release Date: Jan-2023
Expire Date: Apr-2024
Application: NC-200 PQ Kit 4 is part of the NucleoCounter® system, and is used for performance qualification (PQ) using the Via1-Cassette™ and the NucleoCounter NC-200™. For further information please refer to package insert (992-0031) and technical notes (994-0222 and 994-0223).

Lot no. 0123-211		Test Results Master Instrument	Acceptance Criteria Customer Instrument	
			Acceptance Range	Variation Limits
Beads LOW (Low Count / High Viability) No. 071-0024	Total Count (Beads/ml)	6.60×10^4	$4.62 - 8.57 \times 10^4$	CV% < 28%**
	Viability (%)	99	93 - 100	CV% < 3%
Beads MEDIUM (Medium Count / Medium Viability) No. 071-0025	Total Count (Beads/ml)	4.85×10^5	$3.64 - 6.06 \times 10^5$	CV% < 11%**
	Viability (%)	52	45 - 59	CV% < 5%
Beads HIGH (High Count / Low Viability) No. 071-0026	Total Count (Beads/ml)	3.93×10^6	$3.22 - 4.64 \times 10^6$	CV% < 6%
	Viability (%)	0	0 - 6	SD < 2*

* = When viability is close to 0, even insignificant changes in dead counts can result in extremely high CV%. Instead, Standard Deviation (SD) is used to evaluate the result.

** = The high Coefficient of Variation is primarely due to Poisson distribution .

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets or exceeds all product specifications.

This lot was released by:

Camilla Poulsen, QC Specialist

Doc. No.: 992-0032 v. 1.0
ChemoMetec A/S • Gydevang 43 • DK-3450 Allerod • Denmark • support@chemometec.com • www.chemometec.com