Certificate of Analysis for lot: 0223-207

NC-200 PQ Kit

Using the Via1-Cassette™ and Viability and Cell Count Assay on the NucleoCounter® NC-200™

Part number: 912-0014
Lot number: 0223-207
Release date: Feb-2023
Exp. Date: May-2024
Application: **NC-200 PQ Kit** is part of the NucleoCounter® NC-200™ system and is used for performing performance qualification (PQ) using the **Via1-Cassette™ and the NucleoView® Software on NC-200™**. For further information please refer to package insert.

	Results on Master Instrument	Accepted Range	CV %
Beads Low P/N: 071-0032	Count **0.38 x 10^6 beads/mL**	**0.32 - 0.45 x 10^6 beads/mL**	< 10 %
	Viability **94.8 %**	91 - 97 %	< 3 %
Beads Medium P/N: 071-0033	Count **0.96 x 10^6 beads/mL**	**0.82 - 1.11 x 10^6 beads/mL**	< 8 %
	Viability **94.9 %**	91 - 97 %	< 3 %
Beads High P/N: 071-0034	Count **1.87 x 10^6 beads/mL**	**1.59 - 2.15 x 10^6 beads/mL**	< 6 %
	Viability **94.6** %	91 - 97 %	< 3 %

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets or exceeds all product specifications.

This lot was released by: Camilla Poulsen, QC Specialist