Certificate of Analysis for lot: 1022-205

NC-200 PQ Kit

Using the Via1-Cassette™ and Viability and Cell Count Assay on the NucleoCounter® NC-200™

Part number: 912-0014
Lot number: 1022-205
Release date: Oct-2022
Exp. Date: Jan-2024
Application: **NC-200 PQ Kit** is part of the NucleoCounter® NC-200™ system and is used for performing performance qualification (PQ) using the **Via1-Cassette™ and the NucleoView® Software on NC-200™**. For further information please refer to package insert.

	Results on Master Instrument	Accepted Range	CV %
Beads Low P/N: 071-0032	Count **0.36 x 10^6 beads/mL**	**0.29 - 0.42 x 10^6 beads/mL**	< 10 %
	Viability **94.7 %**	91 - 97 %	< 3 %
Beads Medium P/N: 071-0033	Count **0.93 x 10^6 beads/mL**	**0.79 - 1.07 x 10^6 beads/mL**	< 8 %
	Viability **94.8 %**	91 - 97 %	< 3 %
Beads High P/N: 071-0034	Count **1.83 x 10^6 beads/mL**	**1.55 - 2.10 x 10^6 beads/mL**	< 6 %
	Viability **94.9** %	91 - 97 %	< 3 %

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets or exceeds all product specifications.

This lot was released by: Camilla Poulsen, QC Assistant