Certificate of Analysis for lot: 0223-204

NC-3000 PQ Kit /Cassette

Using the Via1-Cassette™ and Viability and Cell Count Assay on the NucleoCounter® NC-3000™

Part number: 912-0020
Lot number: 0223-204
Release date: Feb-2023
Exp. Date: May-2024
Application: **NC-3000 PQ Kit /Cassette** is part of the NucleoCounter® system and is used for performing performance qualification (PQ) using the **Via1-Cassette™ and the NucleoView® Software on NC-3000™**. For further information please refer to package insert.

	Results on Master Instrument	Accepted Range	CV %
Beads Low P/N: 071-0038	Count **0.37 x 10^6 beads/mL**	**0.30 - 0.43 x 10^6 beads/mL**	< 10 %
	Viability **93.7 %**	91 - 97 %	< 3 %
Beads Medium P/N: 071-0039	Count **0.94 x 10^6 beads/mL**	**0.80 - 1.08 x 10^6 beads/mL**	< 8 %
	Viability **94.2 %**	91 - 97 %	< 3 %
Beads High P/N: 071-0040	Count **1.83 x 10^6 beads/mL**	**1.56 - 2.10 x 10^6 beads/mL**	< 6 %
	Viability **94.4 %**	91 - 97 %	< 3 %

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets or exceeds all product specifications.

This lot was released by:

Camilla Poulsen, QC Specialist

Template for Certificate of Analysis - Kits
Doc. No.: 880-0011-79F v.1.0 · Issue Date: 22-Mar-2021
ChemoMetec A/S · Gydevang 43 · DK-3450 Allerod · Denmark · Support@chemometec.com · www.chemometec.com