NC-250 PQ Kit

Using the NC-Slide™ A2 or A8 and Viability and Cell Count Assay on the NucleoCounter® NC-250™

Product No.: 912-0024
Lot No.: 0523-203
Release Date: May-2023
Expiry Date: Aug-2024
Application: **NC-250 PQ Kit** is part of the NucleoCounter® NC-250™ system and is used for performing performance qualification (PQ) using the **NC-Slide™ A2 or A8 and the NucleoView® Software on NC-250™**. For further information please refer to package

	Results on Master Instrument	Accepted Range	CV %
Beads Low P/N: 071-0045	Count		
	3.89 x 10^5 beads/mL	**2.92 - 4.87 x 10^5 beads/mL**	< 10.0 %
	Viability		
	95.0 %	91.0 - 97.0 %	< 3.0 %
Beads Medium P/N: 071-0046	Count		
	0.97 x 10^6 beads/mL	**0.73 - 1.21 x 10^6 beads/mL**	< 8.0 %
	Viability		
	94.9 %	91.0 - 97.0 %	< 3.0 %
Beads High P/N: 071-0047	Count		
	1.90 x 10^6 beads/mL	**1.42 - 2.37 x 10^6 beads/mL**	< 6.0 %
	Viability		
	95.1 %	91.0 - 97.0 %	< 3.0 %

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets or exceeds all product specifications.

This lot was released by: QC Specialist Camila Poulsen