chemometec

NC-3000 PQ Kit /Cassette

Using the Via1-Cassette™ and Viability and Cell Count Assay on the NucleoCounter® NC-3000™

Product No.: 912-0020
Lot No.: 1123-202
Release Date: Nov-2023
Expiry Date: Feb-2025
Application: **NC-3000 PQ Kit /Cassette** is part of the NucleoCounter® system and is used for performing performance qualification (PQ) using the **Via1-Cassette™ and the NucleoView® Software on NC-3000™**. For further information please refer to package insert.

	Results on Master Instrument	Accepted Range	CV %
Beads Low P/N: 071-0038	Count **3.54 x 10^5 beads/mL**	**2.91 - 4.18 x 10^5 beads/mL**	< 10.0 %
	Viability **94.6 %**	91.0 - 97.0 %	< 3.0 %
Beads Medium P/N: 071-0039	Count **0.91 x 10^6 beads/mL**	**0.77 - 1.04 x 10^6 beads/mL**	< 8.0 %
	Viability **94.8 %**	91.0 - 97.0 %	< 3.0 %
Beads High P/N: 071-0040	Count **1.84 x 10^6 beads/mL**	**1.56 - 2.11 x 10^6 beads/mL**	< 6.0 %
	Viability **94.9** %	91.0 - 97.0 %	< 3.0 %

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets or exceeds all product specifications.

This lot was released by: Jennifer Fermina Madsen, QC Specialist