chemometec

NC-3000 PQ Kit /Slide

Using the NC-Slide™ A2 or A8 and Viability and Cell Count Assay on the NucleoCounter® NC-3000™

Product No.: 912-0019
Lot No.: 0623-212
Release Date: Jun-2023
Expiry Date: Sep-2024
Application: **NC-3000 PQ Kit /Slide** is part of the NucleoCounter® system and is used for performing performance qualification (PQ) using the **NC-Slide™ A2 or A8 and the NucleoView® Software on NC-3000™**. For further information please refer to package insert.

	Results on Master Instrument	Accepted Range	CV %
Beads Low P/N: 071-0035	Count **3.87 x 10^5 beads/mL**	**2.91 - 4.84 x 10^5 beads/mL**	< 10.0 %
	Viability **95.2 %**	91.0 - 97.0 %	< 3.0 %
Beads Medium P/N: 071-0036	Count **0.97 x 10^6 beads/mL**	**0.73 - 1.21 x 10^6 beads/mL**	< 8.0 %
	Viability **94.9 %**	91.0 - 97.0 %	< 3.0 %
Beads High P/N: 071-0037	Count **1.92 x 10^6 beads/mL**	**1.44 - 2.40 x 10^6 beads/mL**	< 6.0 %
	Viability **94.9** %	91.0 - 97.0 %	< 3.0 %

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets or exceeds all product specifications.

This lot was released by: Jennifer Fermina Madsen, QC Specialist