CERTIFICATE OF ANALYSIS

chemometec

XM30 SST Kit 1

Using Count & Viability Protocol on the XcytoMatic® 30™

Product No. 912-0031
Lot No.: 0925-208
Release Date: Sep-2025
Expiry Date: Dec-2026
Application: **XM30 SST Kit 1** is part of the XcytoMatic® XM30™ system and is used for system suitability test (SST) using **the XM-View® Software on XM30™**. For further information please refer to package insert.

	Results on Master Instrument	Accepted Range	CV %
Fluorescent beads P/N: 071-0056	Count **1.07E+06 beads/mL**	**9.06E+05 - 1.23E+06 beads/mL**	CV % < 10.0 %
	Viability **49.2 %**	45.0 % - 55.0 %	CV % < 5.0 %
Negative control P/N: 071-0057	Count **1.62E+02 beads/mL**	**0.00E+00 - 2.00E+04 beads/mL**	NA
Carryover	Carryover Limit (%) **< 1.0 %**		

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets all product specifications.

This lot was released by: Nikita Rosted Christensen, QC Lead