CERTIFICATE OF ANALYSIS

chemometec

NC-200 PQ Kit

Using the Via1-Cassette™ and Viability and Cell Count Assay on the NucleoCounter® NC-200™

Product No.: 912-0014
Lot No.: 1125-206
Release Date: Nov-2025
Expiry Date: Feb-2027
Application: NC-200 PQ Kit is part of the NucleoCounter® NC-200™ system and is used for performing performance qualification (PQ) using the **Via1-Cassette™ and the NucleoView® Software on NC-200™**. For further information please refer to package insert.

	Results on Master Instrument	Accepted Range	CV %
Beads Low P/N: 071-0032	Count **3.77 x 10^5 beads/mL**	**3.09 - 4.45 x 10^5 beads/mL**	< 10.0 %
	Viability **95.1 %**	91.0 - 97.0 %	< 3.0 %
Beads Medium P/N: 071-0033	Count **1.06 x 10^6 beads/mL**	**0.90 - 1.22 x 10^6 beads/mL**	< 8.0 %
	Viability **95.0 %**	91.0 - 97.0 %	< 3.0 %
Beads High P/N: 071-0034	Count **2.08 x 10^6 beads/mL**	**1.76 - 2.39 x 10^6 beads/mL**	< 6.0 %
	Viability **94.9 %**	91.0 - 97.0 %	< 3.0 %

This lot has been tested according to the internal ChemoMetec quality procedures. The lot meets all product specifications.

This lot was released by: Nikita Rosted Christensen, QC Lead